



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

AEROMECH EQUIPMENTS PRIVATE LIMITED, 3/1760,F5, SRISUDHARSAN FLAT,  
BELL NAGAR, CHENNAI, TAMIL NADU, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3329

**Page No**

1 of 2

**Validity**

15/12/2021 to 14/12/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	MECHANICAL-PRESSURE INDICATING DEVICES	Pneumatic Pressure: Low Pressure# Manometer, Magnehelic Gauge, Low Pressure Gauge Low Pressure Calibrator.	Using Digital Manometer by Comparison Method based on DKD-R 6-1	-24.5 mbar to 24.5 mbar	0.17mbar



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

AEROMECH EQUIPMENTS PRIVATE LIMITED, 3/1760,F5, SRISUDHARSAN FLAT,  
BELL NAGAR, CHENNAI, TAMIL NADU, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3329

**Page No**

2 of 2

**Validity**

15/12/2021 to 14/12/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Site Facility					
1	MECHANICAL-PRESSURE INDICATING DEVICES	Pneumatic Pressure: Low Pressure# Manometer, Magnehelic Gauge, Low Pressure Gauge Low Pressure Calibrator.	Using Digital Manometer by Comparison Method based on DKD-R 6-1	-24.5 mbar to 24.5 mbar	0.17mbar

\* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.